


GRADUATED PIPETTES

GRADUATED PIPETTE, FOR COMPLETE DELIVERY, WITH SAFETY BULB, QUALICOLOR, COLOR CODE, CLASS AS




Meßpipette, für vollständigen Auslauf, mit Sicherheitskugel, QUALICOLOR, COLOR CODE, Klasse AS
Pipeta dělená na úplný výtok s bezpečnostní baničkou, QUALICOLOR, COLOR CODE, třída AS

ISO 835						
Code	Capacity (ml)	Subdivision (ml)	Color code	Permitted error (\pm ml)	Drawing	Packing (pcs)
632 434 156 312	0.5	0.01	2x yellow	0.006	1679/b	50
632 434 156 314	1	0.01	yellow	0.007	1721	50
632 434 156 516	2	0.02	black	0.010	1722	50
632 434 156 618	5	0.05	red	0.030	1723	50
632 434 156 719	10	0.1	orange	0.050	1724	50
632 434 156 722	20 *	0.1	2x yellow	0.100	1725	25
632 434 156 723	25	0.1	white	0.100	1726	25
632 434 156 725	50 *	0.1	2x orange	0.200	1727	5
632 434 156 831	100 *	0.2	red	0.200	1728	5

Notice: * not according to ISO standard Soda-lime glass

GRADUATED PIPETTE, FOR COMPLETE DELIVERY, WITH SAFETY BULB, QUALICOLOR, COLOR CODE, CLASS B




Meßpipette, für vollständigen Auslauf, mit Sicherheitskugel, QUALICOLOR, COLOR CODE, Klasse B
Pipeta dělená na úplný výtok s bezpečnostní baničkou, QUALICOLOR, COLOR CODE, třída B

ISO 835						
Code	Capacity (ml)	Subdivision (ml)	Color code	Permitted error (\pm ml)	Drawing	Packing (pcs)
632 434 066 312	0.5	0.01	2x yellow	0.010	1679/a	50
632 434 066 714	1	0.1	red	0.010	1702/a	50
632 434 066 314	1	0.01	yellow	0.010	1702	50
632 434 066 716	2	0.1	green	0.020	1703/a	50
632 434 066 316	2 *	0.01	2x white	0.020	1703/b	50
632 434 066 516	2	0.02	black	0.020	1703	50
632 434 066 718	5	0.1	blue	0.050	1704/a	50
632 434 066 618	5	0.05	red	0.050	1704	50
632 434 066 719	10	0.1	orange	0.100	1705	50
632 434 066 722	20 *	0.1	2x yellow	0.200	1706	25
632 434 066 723	25	0.1	white	0.200	1707	25
632 434 066 725	50 *	0.1	2x orange	0.400	1708	5
632 434 066 831	100 *	0.2	red	0.400	1709	5

Notice: * not according to ISO standard Soda-lime glass

25 pirkimo dalis

GRADUATED PIPETTE, FOR PARTIAL DELIVERY, QUALICOLOR, COLOR CODE, CLASS AS



Meßpipette, für teilweisen Auslauf, QUALICOLOR, COLOR CODE, Klasse AS
Pipeta dělená na částečný výtok, QUALICOLOR, COLOR CODE, třída AS

ISO 835						
Code	Capacity (ml)	Subdivision (ml)	Color code	Permitted error (\pm ml)	Drawing	Packing (pcs)
632 434 335 308	0.1	0.01	white	0.006	6440	50
632 434 335 310	0.2	0.01	black	0.006	6441	50
632 434 335 312	0.5	0.01	2x yellow	0.006	6443	50
632 434 335 714	1	0.1	red	0.007	3854	50
632 434 335 314	1	0.01	yellow	0.007	3633	50
632 434 335 716	2	0.1	green	0.010	3855	50
632 434 335 316	2 *	0.01	2x white	0.010	3856	50
632 434 335 516	2	0.02	black	0.010	3634	50
632 434 335 718	5	0.1	blue	0.030	3857	50
632 434 335 618	5	0.05	red	0.030	3635	50
632 434 335 719	10	0.1	orange	0.050	3636	50
632 434 335 722	20 *	0.1	2x yellow	0.100	3637	25
632 434 335 723	25	0.1	white	0.100	3638	25
632 434 335 825	50 *	0.2	black	0.200	3639	5
632 434 335 831	100 *	0.2	red	0.200	3640	5

Notice: * not according to ISO standard Soda-lime glass